

Location: _____

Refrigerator/Freezer Temperature Chart

Month/Year _____

Refrigerator range: 35 to 46°F or 2 to 8°C

Freezer range: less than 5°F or less than -15°C

REFRIGERATOR

FREEZER

Temp		DATE																																	
		1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31			
15	59																																		
14	57																																		
13	55.5																																		
12	54																																		
11	52																																		
10	50																																		
9	48																																		
8	46																																		
7	44.5																																		
6	43																																		
5	41																																		
4	39																																		
3	37.5																																		
2	36																																		
1	34																																		
0	32																																		
-1	30																																		
-2	28																																		
-3	27.5																																		
-4	25																																		
-5	23																																		
-6	21																																		
-7	19																																		
-8	17.5																																		
-9	16																																		
-10	14																																		
-11	12																																		
-12	10.5																																		
-13	9																																		
-14	7																																		
-15	5																																		
-16	3																																		
-17	1																																		
-18	-0.5																																		
-19	-2																																		
-20	-4																																		
-21	-6																																		
-22	-8																																		
INITIALS AM																																			
INITIALS PM																																			

Fill out corrective action in the lines below if temperature reading falls in any shaded areas.

Refrigerator/freezer temperatures shall be monitored and recorded daily unless used for storage of vaccines. Vaccine refrigerators are recorded twice daily. Place a dot (*) for the refrigerator and the freezer readings in the box nearest the temperature for the date indicated. Document on the lines below, continue on the back of the page, any out of range readings below and indicate the corrective action taken. Consult the duty pharmacist or the immunization program manager regarding stability/usability of drugs if out of range. Areas not staffed 7 days per week must use a hi/low thermometer or other method and document history reading for days not staffed.

Date: _____ Action taken: _____
